

10-15-97

Standardize

pH 4 buffer check - 4.006 (7.22) AM

pH 10 buffer check - 10.013 (7.29) AM

Slope/ess. - 100.1

pH 7 buffer check - 6.979

(07:38) AM

Blank non-washed - 7.695

(07:45) AM

pH 7 buffer check - 6.977

(07:52) AM

Blank (Cig) - 6.936

(08:24) AM

pH 7 buffer check - 6.975

(08:30) AM

Im#16 - 5.326

(08:38) AM

pH 7 buffer check - 6.975

(08:44) AM

2 RIF (un-taped) #1 - 5.250

(09:00) AM

pH 7 buffer ^{check} ~~check~~ - 6.975

(09:06) AM

2 RIF (un-taped) #2 - 5.324

(09:23) AM

pH 7 buffer check - 6.973

(09:29) AM

2 RIF (un-taped) #3 - 5.268

(09:40) AM

pH 7 buffer check - 6.971

(09:46) AM

2 RIF (un-taped) #4 - 5.281

(09:54) AM

pH 7 buffer check - 6.972

(09:59) AM

2 RIF (un-taped) #5 - 5.280

(10:08) AM

pH 7 buffer check - 6.972

(10:14) AM

1 R3F (un-taped) #1 - 5.266

(10:25) AM

pH 7 - buffer check - 6.972

(10:31) AM

1 R3F (un-taped) #2 - 5.244

(10:42) AM

pH 7 buffer check - 6.972

(10:48) AM

1 R3F (un-taped) #3 - 5.240

(10:56) AM

pH 7 buffer check - 6.977

(11:02) AM

1 R3F (un-taped) #4 - 5.224

(11:17) AM

pH 7 buffer check - 6.970

(11:22) AM

PM3000375872

1 R3 F (un-taped) #5 - 5.221

PH 7 buffer check - 6.970

IM #16 - 5.310

PH 7 buffer check - 6.971

1 R4 F (un-taped) #1 5.287

PH 7 buffer check - 6.968

1 R4 F (un-taped) #2 - 5.177

PH 7 buffer check - 6.969

1 R4 F (un-taped) #3 - 5.329

PH 7 buffer check - 6.967

1 R4 F (un-taped) #4 - 5.323

PH 7 buffer check - 6.969

1 R4 F (un-taped) #5 - 5.276

PH 7 buffer check - 6.967

1 R5 F (un-taped) #1 -

PH 7 buffer check

1 R5 F (un-taped) #2 -

PH 7 buffer check -

1 R5 F (un-taped) #3 -

PH 7 buffer check -

1 R5 F (un-taped) #4 -

PH 7 buffer check -

1 R5 F (un-taped) #5 -

PH 7 buffer check -

(11:28) AM

(11:34) AM

(11:44) AM

(11:50) AM

(11:56) AM

(12:02) PM

(12:15) PM

(12:21) PM

(12:31) PM

(12:37) PM

(12:50) PM

(12:56) PM

(01:04) PM

(01:10) PM